Test Lab Work Request Form

Date Submitted: 10 March, 2000 Tracking #: TLW 00100
Project #: 241095 Engineer: J.R.SNEDEKER

Test Objective:

TLW0010O - Weight of Bolt Assembly:

The bolt assembly, disassembled from the rifle, will be weighed. The test samples will be weighed on the Mettler Toledo digital balance (PB8000) located in the Metrology Lab. The rifle's bolt assembly will be weighed once each.

Test Description:

Method:

- Check to be sure that the bolt is correctly tagged with the last four digits of its rifle's serial number. This bolt assembly must be returned to its original rifle or the headspace may change.
- Clean the platen of the digital balance, if necessary.
- If the balance is not already on and has been turned on at least 30 mirrutes for warm-up, turn the balance on and wait 30 minutes for the balance circuitry to stabilize.
- Run the balance calibration routine if necessary
- Make sure the units are set to "lb."
- Carefully place the bolt assembly with the bolt assembly's approximate front to rear balance point directly over the center of the balance place.
- When the scale settles down, record the weight in lb. to the nearest 0.1 lb. (Note that the scale has three decimal points displayed.)
- Re-assemble the bolt on its corresponding barreled action.

Data Required:

- Rifle serial number
- Weight to the nearest 0.1-lb.
- Serial number of the Mettler PB8000 balance (it should be SN 2114475246)

Resource Usage:	Test Results Required:					
Manpower Requirements -	Formal Report: Data Only: X REQUESTED Completion Date:					
Facility Requirements -						
Required Materials/Parts/Equipment (include quantities):						
Test Parts Availability Date:						
Start Date: 12 Sept 2000	Test Assigned To:					
Completion Date: 12 Sept 2000 Report Date:	U. Cawent					

Weight of Bolt Assembly - TLW 00100 weight of Barrel Assembly - TLW 0010N weight of Stock - TLW 0010M Overall weight - TLW 0010L

1	Overall	B01+	Barrel	Stock	Take Down
B1 -	6.916*	.654	3.867	2.354	.042 1074
,BS -	6.881*	.653	3.841	2.345	.042 1346
B3 -	6.891*	,655	3.844	2.351	. 047 [10]
B4 -	6,893×	, 654	3.853	2.344	.042 138
85 -	6.899*	,655	3.857	2.346	042 1340
B6 -	6.905*	.654	₹ 863	2,347	,042 1328
B7 -	6.897*	.654	3.857	2.344	.042 [1333]
B8 -	6.874 ×	·. 6\$3	8.8 3 7	2.343	.042 1083
B9 -	6889*	.65 ⁴	3.856	2.338	,042 [1063]
\$10 -	6.400¥	.625.	3.861	2.343	.042 [1069]

Magazine = .215 - Had only one magazine, others taken for testing.

*Overall weight does not include the magazine weight.

All measurements in Lbs.